Probe thermometer check log

Business _____ Month _____ Year _____

Iced water: fill a cup with crushed ice and a little cold water, wait a minute, probe the middle. The reading should be between -1 °C and +1 °C.

Boiling water: probe water at a rolling boil. The reading should be between 99 °C and 101 °C. Outside these ranges, replace the probe or have it calibrated.

Date	Thermometer	Iced water °C	Boiling °C	OK?	Action	Initials